



IAPMO Plumbing Codes and Standards India Pvt. Ltd.

No.43, 4th Floor, PMR Towers, Beratena Agrahara, Near Hosa Road junction, Bangalore, Karnataka 560100



CERTIFICATE OF LISTING

IAPMO Plumbing Codes and Standards India Private Ltd. (aka IAPMO India) is a product certification body which lists and inspects samples taken from the supplier's stock or from the market or a combination of both to verify compliance to the requirements of applicable codes and standards. This activity is coupled with periodic surveillance of the supplier's factory and warehouses as well as the assessment of the supplier's Quality Assurance System. This listing is subject to the conditions set forth in the characteristics below and is not to be construed as any recommendation, assurance or guarantee by IAPMO India of the product acceptance by Authorities Having Jurisdiction.

Licensee: ROCKFOREST TECHNOLOGIES INDIA PRIVATE LIMITED

Address: No. 110, 4th Main Road, St. Ann's School Road, Akshayanagar, Bangalore | Karnataka | India.

File No: DM-IN 2001

Product: UV Cabinet

Void After: June 2021

IDENTIFICATION: Each **EI-CLEAiR, Model No: RAS2020** shall be permanently and legibly marked with the Disinfection Mark with manufacturer's name or trademark, production code, a code number identifying the product and the date of manufacture.

CHARACTERISTICS: The UV cabinet is used for decontaminating air borne microflora in closed rooms and surfaces of smaller articles (by keeping them inside the UV cabinet). As per manufacturers specification, the product contains a particulate trapping filter and UV LEDs. Besides, there is fan for recirculating air through the product. Based on test report, the product is found to be effective in reducing air borne total bacterial counts, yeast & mold counts from a room environment with an approximate volume of 2000 cubic feet to the tune of $\geq 56\%$ within a duration of 2 hours and the performance increased to at least 90% reduction both for bacteria and fungi. Also the naturally occurring airborne bacteria and fungi on articles namely spoon, calculator and vegetable cutting board were reduced at least by 91% in 10 minutes of exposure.

Besides the above the product is also found to reduce induced microbial contaminants on various inanimate articles at least by 99.8% for bacteria, MS2 bacteriophage and fungi in 10 minutes of exposure.

Above mentioned product is independently tested by using customized test protocols developed by Aquadiagnostics Water Research & Technology Center Limited and found in compliance with above stated performance data.

PRODUCT DETAILS: EI-CLEAiR, Model No: RAS2020

Dr. K. Chandrasekhar
IAPMO India

GP Russ Chaney
CEO, The IAPMO Group



This listing period is based upon the last date of the month indicated on the Effective Date and Void After Date shown above. Any change in material, manufacturing process, marking or design without having first obtained the approval of IAPMO India, or any evidence of non-compliance with applicable codes and standards or of inferior workmanship, may be deemed sufficient cause for revocation of this listing. Production of or reference to this form for advertising purposes may be made only by specific written permission of IAPMO India. Any unauthorized alteration of this certificate will be grounds for revocation of this listing.